


## Start therapy

**1**




Fit the garment to the patient.


**2**

Connect the garment to the pump.

**3**



Turn the mains power switch to On (I).



If a single garment is attached, press the Single Leg button.

## Adjust pressure


**1**



40 mmHg

Check that the display is showing the desired pressure when the garments are inflated. Check the display at regular intervals.


**2**



Rotate pressure control knob until the desired pressure is displayed.

## Stop therapy

**1**



Turn the mains power switch to off (O).

**2**

Disconnect the garment from the pump.