
# **Shower and Disinfection Panels**




# **Disinfection Flow Control**




Turn the disinfection control to start and adjust disinfection flow.

### **Flow Meter**




Shows mix ratio between disinfectant and water. To adjust, turn the suction part on the bottle cap. Make sure the mix is correct when starting disinfection. (See IFU.)

## **Shower and Disinfection Handle**




Shower handle is grey or white. Disinfection handle is red or yellow. Turn on flow control. Hold the trigger in for constant water flow (A). On shower handle, press lock button for continuous flow (B).

#### **Thermometer**




Displays temperature of outcoming water and disinfectant.

# **Lockable Door for Disinfectant**




The disinfection bottle is behind a locked door. Make sure to keep door locked and key safely stored away.

## **Shower Flow Control**




Turn the shower flow control to start the water and adjust the water flow.

# **Temperature Control**



Turn the temperature control to adjust the water temperature or disinfectant temperature.

## Flush valve or Flush button



If a flush sink is connected, flush it by moving the flushing valve (A) forwards or backwards, or by pushing the flush button (B).